

**MEK-3DL HEMATOLOGY CONTROL 2.0mL**

**ASSAY SHEET**

<b>LOT</b>	B23YL
	2024-02-05
<b>OPEN</b>	14DAYS

**MEK-1301,1302,1305 Barcode Entry**

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420	MEK -6500 -6510	MEK -6450	MEK -6550	MEK -1301 -1302 -1305
<b>WBC</b>	10 <sup>3</sup> /μL	±0.6	2.0	2.0	2.0	2.0	1.98
<b>RBC</b>	10 <sup>6</sup> /μL	±0.15	2.38	2.38	2.38	2.38	2.32
<b>HGB</b>	g/dL	±0.4	5.9	6.1	5.9	6.1	5.95
<b>HCT</b>	%	±2.0	17.5	17.5	17.5	17.5	19.5
<b>MCV</b>	fL	±9.0	73.6	73.6	73.6	73.6	84.2
<b>MCH</b>	pg	±2.4	24.8	25.6	24.8	25.6	25.6
<b>MCHC</b>	g/dL	±3.0	33.7	34.9	33.7	34.9	30.5
<b>PLT</b>	10 <sup>3</sup> /μL	±30	66	66	66	66	74.8
<b>MPV</b>	fL	±3.0	6.9	8.9	6.9	8.9	8.9
<b>RDW-CV *</b>	%	±3.0	17.6	17.6	17.6	17.6	16.1
<b>LY%</b>	%	±10.0	63.2	63.2	63.2	63.2	65.49
<b>MO%</b>	%	±5.0	8.4	8.4	8.4	8.4	4.11
<b>GR%</b>	%	±10.0	28.4	28.4	28.4	28.4	30.40
<b>LY</b>	10 <sup>3</sup> /μL	±0.5	1.3	1.3	1.3	1.3	1.30
<b>MO</b>	10 <sup>3</sup> /μL	±0.2	0.2	0.2	0.2	0.2	0.08
<b>GR</b>	10 <sup>3</sup> /μL	±0.5	0.5	0.5	0.5	0.5	0.60

<b>MEK-1305</b>
<b>QC ESR</b>
<b>50±17 mm</b>

<b>Manual Registration</b>	
Lot	23Y
Level	Low
Control	MEK-3D

\* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN



0694-901684N