

MEK-3DH HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B242H
	2024-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1

2

3

		RANGE	MEK -6400 -6410 -6420	MEK -6500 -6510	MEK -6450	MEK -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	20.9	20.9	20.9	20.9	20.47
RBC	10 ⁶ /μL	±0.25	5.87	5.87	5.87	5.87	5.90
HGB	g/dL	±0.7	19.3	19.3	19.3	19.3	19.42
HCT	%	±4.0	54.1	54.1	54.1	54.1	62.6
MCV	fL	±7.0	92.2	92.2	92.2	92.2	106.1
MCH	pg	±2.8	32.9	32.9	32.9	32.9	32.9
MCHC	g/dL	±3.0	35.7	35.7	35.7	35.7	31.0
PLT	10 ³ /μL	±70	592	592	592	592	583.7
MPV	fL	±3.0	5.8	8.1	5.8	8.1	8.5
RDW-CV *	%	±3.0	16.2	16.2	16.2	16.2	17.1
LY%	%	±15.0	19.6	19.6	19.6	19.6	14.75
MO%	%	±8.0	3.3	3.3	3.3	3.3	1.65
GR%	%	±15.0	77.1	77.1	77.1	77.1	83.60
LY	10 ³ /μL	±3.0	4.1	4.1	4.1	4.1	3.02
MO	10 ³ /μL	±1.5	0.7	0.7	0.7	0.7	0.34
GR	10 ³ /μL	±3.0	16.1	16.1	16.1	16.1	17.11

Manual Registration	
Lot	242
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN

MEK-3DH

B242H

2024-05-05

0694-9016860